



Characterization of the ALP Somatic Cell Count Standard (SCCS) – a “Proficiency Testing”

T. Berger, C. Egger, S. Brunner

EU-RL MMP 2011 workshop, anses, Maisons-Alfort, 2. Mai 2011



History of ALP Somatic Cell Count Standard (SCCS)

Beginnings

- approx. 1984 cooperation of ALP and VetSuisse faculty to produce a SCCS (Schällibaum, Limacher, Merminod)
- 1986 extraction of thymocytes and (because of BSE) later leucocytes from blood of young bulls by Bommeli corporation, preparation of ALP SCCS for Swiss raw milk control laboratories; reference value based only on ALP value
- approx. 1995 first customers in other countries
- 2002 more customers in other countries



History of ALP SCCS

Improvements

- 2004 reference value based on an international characterization
- 2004/2005 stability problems, transfer of the cell extraction to ALP, improvement of the procedure (Meyer, Bühlmann, Raemy, Brunner, Aebi)
- 2008 optimization of fixation procedure (Egger, Raemy, Brunner)



History of ALP SCCS

Use of ALP SCCS

- Germany
- Netherlands
- Austria
- Switzerland
- Italy
- Chile





Use, quality, quantities and price

Use and quality

- ALP-SCCS is used for the evaluation of comparability, reproducibility and traceability of automated cell count measurements
- it covers the range from 150'000 to 450'000 cells/mL on two levels
- quality is checked at date of production and within the whole time of shelf life
 - homogeneity control during bottling
 - weekly check of reference values at ALP
 - evaluation of measurement results voluntary sent back by labs (evaluation report)



Use, quality, quantities and price

Quantities and price

- volumes: 20, 27 and 40 mL in plastic tubes
- production 3x/year
- shelf life 6 months
- distribution frequency: monthly
- price: CHF -.45/mL incl. delivery



Characterization of reference value

- based on DMSCC and ASCC measurements
- using a pool of experienced international laboratories:
 - MCC-Vlaanderen, Lier (BE)
 - BfEL/Max-Rubner-Inst, Kiel (DE)
 - mpr-BW, Kirchheim (DE)
 - mpr-BY, Wolnzach (DE)
 - Greek Dairy Org., Ioannina (GR)
 - ICBA, Caesaria (IL)
 - AIA Isl, Roma (IT)
 - LA-Santander, Santander (SP)
 - LACM, Valdivia (CL)
 - Qlip, Zutphen (NL)
 - RIKILT, Wageningen (NL)
 - mpr-OÖ, Ried (AT)
 - University of Vienna/Vet. Dep. (AT)
 - LNIV/ALIP, Lisboa/Lousada (PT)
 - ALP, Liebefeld (CH)
 - Suisselab AG, Zollikofen (CH)
 - CMIO C.T.LAB, Nicosia (CY)
 - SV Südtirol/FL Alto Adige, Bolzano (IT)



Characterization of reference value

Schweizerische Eidgenossenschaft Confédération suisse Confederazione Svizzera Confederaziun svizra									
Agroscope Liebefeld-Posieux Research Station ALP									
Characterization of SCC reference material									
Name of Laboratory:	LABORATORIO STANDARD LATTE								
Samples arrival date:	03.09.2008	slope:							
State of the samples at arrival:	GOOD	intercept:							
Date of launching of the analyses:	04.09.2008								
Automated measurement									
Type of equipment:	BENTLEY SOMACOUNT 150								
Raw data	Sample low 1	Sample low 2	Sample low 3	Sample high 4	Sample high 5	Sample high 6			
Measurement 1	232	234	219	507	505	518			
Measurement 2	219	229	221	507	520	516			
Measurement 3	229	223	230	498	517	511			
Measurement 4	229	218	225	501	496	511			
Measurement 5	220	225	225	495	510	508	low	high	
Mean value	226	226	224	502	510	513	Total mean	225	508
Standard deviation	6	6	4	5	10	4	Total standard deviation	5	8
Coefficient of variation %	2.6	2.7	1.9	1.1	1.9	0.8	Total CV %	2.3	1.6
Microscopic measurement									
Raw data sample "high"	Operator 1: countings			Operator 2: countings					
	1st	2nd	3rd	1st	2nd	3rd			
Film a	521								
Film b	520						high		
Mean value	173	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	Total mean	521	

- procedure: the labs receive samples by normal post (stability has been checked) and an excel evaluation-sheet by email



Characterization of reference value

- results are returned by email within 14 days
- production and characterization are included in ISO 9001 certification and ISO 17025 accreditation
- no accreditation for PTs according to ISO 17043
- ISO 13528 is not explicitly followed



Characterization of reference value

AGROSCOPE · SCC-REFERENCE-MATERIAL · 10C, September 2010

2 → Statistical data

2.1 → Routine instruments, all data

Results and group means in SCC/μl

Code	Count	L_1	L_2	L_3	L_mean	H_1	H_2	H_3	H_mean
T_01_1	1	284	306	271		557	578	610	
T_01_1	2	281	294	278		568	579	567	
T_01_1	3	282	274	284		581	577	555	
T_01_1	4	292	289	267		561	569	563	
T_01_1	5	304	282	280	284.5	576	564	558	570.9
T_01_2	1	313	295	294		588	605	586	


- statistical treatment of data includes following values [separately for Direct Microscopic SCC (DMSCC) and Automated SCC (ASCC)]:
 - outlier test, overall mean value, lab mean, standard deviation, coefficient of variation, z-score, repeatability, reproducibility, graphs
 - anonymised evaluation of labs




Characterization of reference value

- reference value is the mean value of the AFEMA laboratories → observed reproducibilities are usually better in equilibrated measurement networks


Recovery "high"




Precision "high"



Recovery "low"



Precision "low"



All-routine-instruments, -deviation-statistics

SCC/μl



LEVEL	LABS*	N	MEAN	S _L	S _p	S _R	RSD _L	RSD _p	RSD _R	r	R	R/r
SSC-low	21	315	295.35	16.05	9.83	18.82	5.44	3.33	6.37	27.81	53.27	1.92
SSC-high	21	315	594.17	19.46	14.10	24.04	3.28	2.37	4.05	39.91	68.02	1.70

Laboratories or instruments, respectively. One instrument excluded for the estimation of the variance components of SSC-high (Cochran-outlier, p=0.017)





Characterization of reference value

high



low



Agroscope SCC reference material September 2010

Reference method and routine method

no outliers removed

Reference

Routine