



Schweizerische Eidgenossenschaft
Confédération suisse
Confederazione Svizzera
Confederaziun svizra

Federal Department of Economic Affairs DEA
Agroscope Liebefeld-Posieux Research Station ALP

Production and Characterization of reference material by ALP

Thomas Berger, Sandra Brunner, Lotti Egger

02. October 2008



History of ALP Somatic Cell Count Standard (SCCS)

Beginnings

- approx. 1984 cooperation of ALP and VetSuisse faculty to produce a SCCS (Schällibaum, Limacher, Merminod)
- 1986 extraction of thymocytes and (because of BSE) later leucocytes from blood of young bulls by Bommeli corporation, preparation of ALP SCCS for Swiss raw milk control laboratories; reference value based only on ALP value
- approx. 1995 first customers in other countries
- 2002 more customers in other countries



History of ALP SCCS

Improvements

- 2004 reference value based on an international characterization
- 2004/2005 stability problems, transfer of the cell extraction to ALP, improvement of the procedure (Meyer, Bühlmann, Raemy, Brunner, Aebi)
- 2008 optimization of fixation procedure (Egger, Raemy, Brunner)



History of ALP SCCS

Use of ALP SCCS

- Germany
- Austria
- Switzerland
- Italy
- Croatia
- Chile





Preparation of SCCS

Main steps

- Fresh blood from young bulls containing 10% anticoagulant solution
- Filtration
- Centrifugation, removal of blood serum
- First lysis of red blood cells with EDTA, centrifugation after addition of PBS
- Resuspension of the sediment with PBS
- Second lysis as before
- Centrifugation
- Resuspension of the sediment with PBS
- Third lysis as before
- Centrifugation





Preparation of SCCS

Main steps

- 2 further wash steps with PBS
- Centrifugation
- Resuspension of leucocytes in skim milk containing fixation solution (10% ethanol, Bronopol, and polypropylen-glycol)
- Addition of a defined volume of leucocyte suspension to UHT milk containing Bronopol and polypropylen-glycol to obtain a high and a low standard





Use, quantities and price

- The ALP Somatic Cell Count Standard (SCCS) is used for the evaluation of comparability, reproducibility and traceability of automated cell count measurements
- it covers the range from 150'000 up to 450'000 cells/mL on two levels
- quantities: 7, 20, 27 and 40 mL
- frequency: monthly, bimonthly, quarterly
- Price: CHF -.45/mL incl. postage



Characterization of reference value

- based on DMSCC and ASCC measurements
- using a pool of experienced international laboratories, including:
 - MCC-Vlaanderen, Lier (BE)
 - BfEL/Max-Rubner-Inst, Kiel (DE)
 - mpr-BW, Kirchheim (DE)
 - mpr-BY, Wolnzach (DE)
 - Greek Dairy Org., Ioannina (GR)
 - Greek Dairy Org., Larisa (GR)
 - Greek Dairy Org., Patras (GR)
 - Greek Dairy Org., Pella (GR)
 - ICBA, Caesaria (IL)
 - AIA Isl, Roma (IT)
 - SV Südtirol, Bolzano (IT)



Characterization of reference value

- Qlip, Zutphen (NL)
- RIKILT, Wageningen (NL)
- mpr-OÖ, Ried (AT)
- University of Vienna/
Veterinary Dep. (AT)
- LNIV; Lisboa (PT)
- ALP, Liebefeld (CH)
- Suisselab AG, Zollikofen (CH)
- LA-Santander, Santander (SP)
- CMIO C.T.LAB, Nicosia (CY)

Schweizerische Eidgenossenschaft Confédération suisse Confederazione Svizzera Confederaziun svizra									
Agroscope Liebefeld-Posieux Research Station ALP									
Characterization of SCC reference material									
Name of Laboratory:	LABORATORIO STANDARD LATTE								
Samples arrival date:	03.09.2008	slope:							
State of the samples at arrival:	GOOD	intercept:							
Date of launching of the analyses:	04.09.2008								
Automated measurement									
Type of equipment:	BENTLEY SOMACOUNT 150								
Raw data	Sample low 1	Sample low 2	Sample low 3	Sample high 4	Sample high 5	Sample high 6			
Measurement 1	232	234	219	507	505	518			
Measurement 2	219	229	221	507	520	516			
Measurement 3	229	223	230	498	517	511			
Measurement 4	229	218	225	501	496	511			
Measurement 5	220	225	225	495	510	508			
Mean value	226	226	224	502	510	513	Total mean	low	high
Standard deviation	6	6	4	5	10	4	Total standard deviation	5	8
Coefficient of variation %	2.6	2.7	1.9	1.1	1.9	0.8	Total CV %	2.3	1.6
Microscopic measurement									
Raw data sample "high"	Operator 1: countings			Operator 2: countings					
	1st	2nd	3rd	1st	2nd	3rd			
Film a	521								
Film b	520						high		
Mean value	173	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	Total mean	521	
Standard deviation	16	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	Total standard deviation	1	
Coefficient of variation %	9.4	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	Total CV %	0.1	
Raw data sample "low"	Operator 1: countings			Operator 2: countings					
	1st	2nd	3rd	1st	2nd	3rd			
Film a	161								
Film b	184						low		
Mean value	173	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	Total mean	173	
Standard deviation	16	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	Total standard deviation	16	
Coefficient of variation %	9.4	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	Total CV %	9.4	

- Procedure: the labs receive samples by normal post and an Excel evaluation - sheet by email
- The measurement results are returned by email within 14 days



Characterization of reference value

- Performance of statistical treatment of data including: outlier test, mean value, standard deviation, coefficient of variation, repeatability, reproducibility, Mandel statistics and graphs

A) Routine Instruments, all data, NOT NORMALIZED BY SLOPE-INTERCEPT

Routine Instruments, Results and Group Means in SCC/μl

CODE	COUNT	L 1	L 2	L 3	L MEAN	H 1	H 2	H 3	H MEAN
T_01	1	274	261	268		467	489	488	
T_01	2	266	257	274		492	478	480	
T_01	3	263	276	267		460	482	498	
T_01	4	268	278	266		468	467	465	
T_01	5	274	253	265	267.3	452	475	469	475.3
T_02	1	244	250	254		457	466	477	
T_02	2	250	258	234		458	465	445	
T_02	3	243	254	248		471	453	456	
T_02	4	247	249	249		454	448	453	
T_02	5	238	266	248	248.8	453	436	461	456.9
T_03	1	267	258	247		454	448	478	
T_03	2	248	249	265		458	467	468	
T_03	3	272	258	243		445	451	450	
T_03	4	256	265	264		461	438	462	
T_03	5	257	268	243	257.3	470	456	451	457.1
T_04	1	258	250	247		442	449	426	
T_04	2	256	256	249		437	476	433	

Routine Instruments, Basic Statistics, Instrument Results sorted by Means
Standard ZZS_07C_L, SCC/μl

T_CODE	N	MEAN	SD	CV	Z-VAL	PERCENTS
T_09	15	190.73	12.07	0.06	-2.83	75.17
T_14	15	233.20	7.25	0.03	-0.98	91.91
T_13	15	240.67	32.84	0.14	-0.65	94.85
T_02	15	248.80	7.72	0.03	-0.30	98.06
T_11	15	254.73	5.69	0.02	-0.04	100.39
T_03	15	257.33	9.49	0.04	0.08	101.42
T_04	15	258.13	10.35	0.04	0.11	101.73
T_10	15	258.40	10.32	0.04	0.12	101.84
T_07	15	260.20	10.73	0.04	0.20	102.55
T_15	15	260.73	25.70	0.10	0.23	102.76
T_05	15	265.67	4.67	0.02	0.44	104.70
T_01	15	267.33	7.09	0.03	0.51	105.36
T_06	15	269.67	11.79	0.04	0.62	106.28
T_08	15	272.93	8.56	0.03	0.76	107.57
T_12	5	295.00	7.91	0.03	1.72	116.26



Characterization of reference value

- reference value is the mean value of the AFEMA laboratories → observed reproducibilities are usually better in equilibrated measurement networks
- see ISO/IDF action for a reference system for somatic cell counting

SCC/μl



LEVEL	LABS	N	MEAN	SL	Sr	SR	RSDL	RSDr	RSDR	r	R	R/r
ZZS_07C_L	15	215	253.74	21.34	13.88	25.45	8.41	5.47	10.03	39.27	72.00	1.83
ZZS_07C_H	15	215	461.67	27.43	15.32	31.42	5.94	3.32	6.80	43.32	88.86	2.05

B) Routine Instruments, AFEMA data

SCC/μl

LEVEL	LABS	N	MEAN	SL	Sr	SR	RSDL	RSDr	RSDR	r	R	R/r
ZZS_07C_L	9	150	260.91	5.23	12.02	13.11	2.01	4.61	5.02	33.99	37.08	1.09
ZZS_07C_H	10	150	468.90	10.44	16.35	19.40	2.23	3.49	4.14	46.24	54.87	1.19

AGROSCOPE SCC REFERENCE MATERIAL September 2007
SUMMARY OF OPERATORS PERFORMANCE
STANDARD low AND high COUNTED in DMSCC/μl






Characterization of reference value

AGROSCOPE SCC REFERENCE MATERIAL September 2007


REFERENCE METHOD and ROUTINE METHOD

no outliers removed

high



low




AGROSCOPE SCC REFERENCE MATERIAL September 2007


ROUTINE INSTRUMENTS

no outliers removed, adjusted and normalized values

high



low





Characterization of reference value

